

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001891  
11:01 25/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:05  
STERILIZING 3:05  
136.4 C 225kPa 3:05  
136.4 C 226kPa 3:35  
136.4 C 226kPa 4:05  
136.3 C 224kPa 4:35  
136.4 C 223kPa 5:05  
136.3 C 225kPa 5:35  
136.2 C 224kPa 6:05  
136.1 C 224kPa 6:35  
136.1 C 224kPa 6:35  
VENTING 6:35  
AIR DRYING 7:28  
CYCLE COMPLETE 25:35  
-----

Digital Signature #  
F385F0A17B412ADF  
-----