
Midmark M11 - v1.0.0
Total Cycles: 1570

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 09:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 73.8 F 0.0
4:00 86.7 F 0.0
6:00 128.9 F 0.1
8:00 157.4 F 0.5
10:00 182.1 F 2.4
12:00 192.6 F 5.9
14:00 206.3 F 9.2
16:00 217.6 F 12.7
18:00 224.4 F 6.3
20:00 234.7 F 10.4
22:00 235.2 F 8.4
24:00 243.6 F 12.3
26:00 241.1 F 10.6
28:00 253.8 F 17.0
30:00 265.1 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.1 F 28.2
2:00 272.6 F 28.7
3:00 271.4 F 27.9
4:00 272.4 F 28.4

Min 271.3 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.1 F 0.0
5:00 213.8 F 0.0
10:00 202.2 F 0.0
15:00 201.4 F 0.0
20:00 200.7 F 0.0
25:00 221.9 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:51
CYCLE COMPLETE

01/31/2022 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____