

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2809

Name: RITTER 1  
Sterilizer ID: V2178198  
01/24/2023 08:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 78.8 F 0.0  
2:00 80.4 F 0.0  
4:00 119.2 F 0.0  
6:00 139.5 F 0.0  
8:00 175.3 F 0.4  
10:00 189.0 F 3.7  
12:00 203.1 F 7.1  
14:00 214.4 F 10.5  
16:00 227.7 F 7.8  
18:00 229.4 F 8.2  
20:00 238.3 F 12.4  
22:00 236.6 F 9.1  
24:00 240.2 F 10.4  
26:00 245.6 F 12.8  
28:00 257.1 F 18.9  
30:00 267.2 F 25.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.3 F 28.8  
2:00 271.4 F 28.1  
3:00 272.4 F 28.9  
4:00 271.5 F 28.1

Min 271.2 F 28.0  
Max 272.4 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 198.1 F 0.0  
10:00 204.6 F 0.0  
15:00 198.9 F 0.0  
20:00 205.3 F 0.0  
25:00 222.0 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:28  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:09:29  
CYCLE COMPLETE

01/24/2023 09:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_