

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1390

Name: RITTER 1  
Sterilizer ID: V2178198  
11/23/2021 04:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.1 F 0.0  
2:00 128.7 F 0.0  
4:00 145.9 F 0.1  
6:00 171.4 F 0.9  
8:00 195.5 F 5.0  
10:00 212.0 F 9.6  
12:00 223.2 F 8.1  
14:00 233.3 F 9.5  
16:00 237.3 F 9.4  
18:00 244.1 F 12.5  
20:00 241.8 F 11.1  
22:00 255.5 F 18.0  
24:00 267.2 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.1  
1:00 272.0 F 28.1  
2:00 272.5 F 28.4  
3:00 272.8 F 28.8  
4:00 272.0 F 28.5

Min 271.7 F 28.0  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 170.9 F 0.0  
5:00 221.4 F 0.0  
10:00 206.1 F 0.0  
15:00 207.5 F 0.0  
20:00 203.7 F 0.0  
25:00 203.5 F 0.0  
30:00 209.2 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 25:17  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 01:03:07  
CYCLE COMPLETE

11/23/2021 05:18 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_