
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008426
15:18 19/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:44
STERILIZING 3:56
136.5 C 223kPa 3:56
136.4 C 226kPa 4:26
136.1 C 223kPa 4:56
136.4 C 224kPa 5:26
136.2 C 223kPa 5:56
136.4 C 223kPa 6:26
136.3 C 224kPa 6:56
136.4 C 223kPa 7:26
136.4 C 223kPa 7:26
VENTING 7:26
AIR DRYING 8:17
CYCLE COMPLETE 10:11

Digital Signature #
FD6C0783BFC1595E
