

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1282

Name: RITTER 1  
Sterilizer ID: V2178198  
10/21/2021 06:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.3 F 0.0  
2:00 72.8 F 0.0  
4:00 78.8 F 0.0  
6:00 107.1 F 0.0  
8:00 140.5 F 0.1  
10:00 176.1 F 0.2  
12:00 197.9 F 0.3  
14:00 214.0 F 3.6  
16:00 223.1 F 6.1  
18:00 230.0 F 9.0  
20:00 239.5 F 12.4  
22:00 235.2 F 8.8  
24:00 242.9 F 11.4  
26:00 243.9 F 12.2  
28:00 241.1 F 10.7  
30:00 253.8 F 17.1  
32:00 264.9 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.6 F 28.7  
2:00 271.6 F 28.2  
3:00 272.6 F 28.6  
4:00 271.5 F 28.1  
  
Min 271.5 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
5:00 226.9 F 0.0  
10:00 181.1 F 0.0  
15:00 166.5 F 0.0  
20:00 190.8 F 0.0  
25:00 189.0 F 0.0  
30:00 188.7 F 0.0  
  
Drying Complete

FILLING: 0:59  
HEATING: 33:54  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:11:60  
CYCLE COMPLETE

10/21/2021 08:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_