
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007949
7:14 09/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:52
STERILIZING 3:53
136.5 C 220kPa 3:53
136.4 C 223kPa 4:23
136.4 C 222kPa 4:53
136.3 C 221kPa 5:23
136.2 C 222kPa 5:53
136.4 C 223kPa 6:23
136.3 C 219kPa 6:53
136.1 C 220kPa 7:23
136.1 C 220kPa 7:23
VENTING 7:23
AIR DRYING 8:15
CYCLE COMPLETE 38:15

Digital Signature #
1A4DB1D0B954CCBC
