
Midmark M11 - v1.0.5
Total Cycles: 3000

Name: RITTER2
Sterilizer ID: V2313228
10/12/2023 09:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.6 F 0.0
2:00 157.1 F 2.0
4:00 167.1 F 5.4
6:00 190.9 F 10.0
8:00 224.9 F 6.0
10:00 222.8 F 9.5
12:00 234.8 F 8.8
14:00 240.7 F 12.3
16:00 240.4 F 10.9
18:00 253.5 F 17.5
20:00 265.4 F 24.6
22:00 272.8 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.5 F 29.5
2:00 272.5 F 30.0
3:00 271.5 F 29.2
4:00 272.4 F 29.9

Min 271.4 F 29.2
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 225.7 F 0.0
10:00 190.2 F 0.0
15:00 186.4 F 0.0
20:00 182.7 F 0.0
25:00 180.2 F 0.0
30:00 177.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

10/12/2023 10:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____