
Midmark M11 - v1.0.0
Total Cycles: 1236

Name: RITTER 1
Sterilizer ID: V2178198
10/06/2021 01:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 89.7 F 0.0
2:00 91.2 F 0.0
4:00 104.8 F 0.0
6:00 133.2 F 0.1
8:00 152.6 F 1.0
10:00 184.2 F 3.7
12:00 205.9 F 6.7
14:00 220.8 F 10.5
16:00 227.8 F 7.6
18:00 232.2 F 9.2
20:00 240.3 F 11.6
22:00 240.6 F 11.1
24:00 236.7 F 8.8
26:00 250.5 F 15.2
28:00 261.2 F 21.3
30:00 271.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 28.9
2:00 271.9 F 28.2
3:00 271.8 F 28.5
4:00 272.7 F 28.8

Min 271.3 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 217.9 F 0.0
10:00 222.7 F 0.0
15:00 219.1 F 0.0
20:00 215.9 F 0.0
25:00 218.8 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:45
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:08:54
CYCLE COMPLETE

10/06/2021 02:37 PM
ACCEPTED: X
REJECTED:
Operator: U01