

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000412  
14:14 26/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:06  
PRESSURIZING 3:08  
STERILIZING 4:09  
136.0 C 228kPa 4:09  
136.4 C 231kPa 4:39  
136.2 C 232kPa 5:09  
136.2 C 231kPa 5:39  
136.3 C 232kPa 6:09  
136.2 C 232kPa 6:39  
136.2 C 231kPa 7:09  
136.4 C 233kPa 7:39  
136.4 C 233kPa 7:39  
VENTING 7:39  
AIR DRYING 8:34  
CYCLE COMPLETE 8:49  
-----

Digital Signature #  
0DC206C3CD7D8AD3  
-----