
Midmark M11 - v1.0.0
Total Cycles: 1350

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 07:15 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	72.4 F	0.0
2:00	73.2 F	0.0
4:00	82.2 F	0.0
6:00	115.9 F	0.0
8:00	144.9 F	0.1
10:00	173.8 F	0.4
12:00	187.8 F	3.1
14:00	201.7 F	5.7
16:00	217.7 F	8.6
18:00	227.2 F	11.9
20:00	229.2 F	6.3
22:00	236.7 F	10.8
24:00	238.1 F	8.2
26:00	245.1 F	13.0
28:00	242.7 F	11.4
30:00	254.3 F	17.3
32:00	264.6 F	23.4
34:00	271.5 F	28.0

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	28.1
1:00	272.3 F	28.5
2:00	272.0 F	28.5
3:00	271.9 F	28.2
4:00	272.3 F	28.8

Min 271.5 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	211.9 F	0.0
	5:00	206.3 F	0.0
	10:00	210.7 F	0.0
	15:00	212.7 F	0.0
	20:00	208.1 F	0.0
	25:00	208.2 F	0.0
Drying Complete	30:00	207.9 F	0.0

FILLING: 0:58
HEATING: 34:02
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:12:17
CYCLE COMPLETE

11/10/2021 08:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____