
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
2.6uS / 1.6ppm
CYCLE NUMBER 000030
13:04 30/09/2019
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 2:44
CYCLE FAULT 15
NOT STERILE
END TIME 2:44

Digital Signature #
1BD4CD1DD1B5FF3D
