

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 006808  
14:20 04/01/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:25  
STERILIZING 3:31  
136.4 C 220kPa 3:31  
136.4 C 223kPa 4:01  
136.2 C 223kPa 4:31  
136.4 C 222kPa 5:01  
136.4 C 223kPa 5:31  
136.4 C 225kPa 6:01  
136.4 C 224kPa 6:31  
136.2 C 223kPa 7:01  
136.2 C 223kPa 7:01  
VENTING 7:01  
AIR DRYING 7:56  
CYCLE COMPLETE 37:56

-----  
Digital Signature #  
677E35C4155E8619  
-----