

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2669

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2022 03:08 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.9 F 0.0  
2:00 164.4 F 1.8  
4:00 170.9 F 5.5  
6:00 204.5 F 10.2  
8:00 222.5 F 4.8  
10:00 231.1 F 9.6  
12:00 239.0 F 9.2  
14:00 240.8 F 12.1  
16:00 239.7 F 10.6  
18:00 253.5 F 17.4  
20:00 265.4 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.6 F 29.3  
2:00 271.7 F 28.5  
3:00 272.6 F 29.4  
4:00 272.3 F 29.0  
  
Min 271.5 F 28.5  
Max 272.7 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 201.8 F 0.0  
10:00 217.2 F 0.0  
15:00 222.7 F 0.0  
20:00 224.5 F 0.0  
25:00 215.1 F 0.0  
30:00 210.0 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 21:45  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:18  
CYCLE COMPLETE

12/13/2022 04:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_