
Midmark M11 - v1.0.0
Total Cycles: 3199

Name: RITTER 1
Sterilizer ID: V2178198
05/04/2023 07:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 85.8 F 0.0
4:00 88.4 F 0.0
6:00 115.7 F 0.3
8:00 149.5 F 1.7
10:00 175.7 F 3.8
12:00 192.7 F 7.0
14:00 202.8 F 10.2
16:00 218.8 F 9.7
18:00 223.3 F 7.5
20:00 234.0 F 11.6
22:00 232.0 F 7.7
24:00 244.7 F 13.2
26:00 242.0 F 11.3
28:00 253.7 F 17.2
30:00 264.1 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.5 F 29.0
2:00 271.5 F 28.1
3:00 272.6 F 28.9
4:00 271.7 F 28.2

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.5 F 0.0
5:00 207.9 F 0.0
10:00 214.9 F 0.0
15:00 227.9 F 0.0
20:00 227.6 F 0.0
25:00 225.4 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:54
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:09:59
CYCLE COMPLETE

05/04/2023 08:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____