
Midmark M11 - v1.0.0
Total Cycles: 1299

Name: RITTER 1
Sterilizer ID: V2178198
10/26/2021 06:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 75.9 F 0.0
4:00 82.8 F 0.0
6:00 121.4 F 0.0
8:00 145.7 F 0.1
10:00 171.6 F 0.1
12:00 198.6 F 0.3
14:00 213.9 F 3.2
16:00 221.7 F 5.6
18:00 229.7 F 8.5
20:00 237.7 F 12.0
22:00 234.0 F 8.3
24:00 245.7 F 13.6
26:00 243.0 F 11.7
28:00 240.1 F 10.3
30:00 253.1 F 16.7
32:00 264.1 F 23.1
34:00 271.7 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.8 F 28.8
2:00 271.5 F 28.0
3:00 272.8 F 28.9
4:00 271.8 F 28.1

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 209.7 F 0.0
10:00 217.7 F 0.0
15:00 209.2 F 0.0
20:00 209.3 F 0.0
25:00 209.1 F 0.0
30:00 204.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 34:07
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:12:13
CYCLE COMPLETE

10/26/2021 08:02 AM
ACCEPTED: X
REJECTED:
Operator: U01