
Midmark M11 - v1.0.5
Total Cycles: 333

Name: RITTER2
Sterilizer ID: V2313228
09/02/2021 02:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 160.6 F 2.8
4:00 178.2 F 6.9
6:00 202.5 F 11.6
8:00 219.3 F 6.5
10:00 230.8 F 12.0
12:00 234.7 F 9.5
14:00 237.1 F 9.5
16:00 248.0 F 14.8
18:00 259.9 F 21.5
20:00 270.5 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.3 F 29.6
2:00 271.4 F 29.1
3:00 272.3 F 29.7
4:00 271.7 F 29.1

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 237.3 F 0.0
10:00 225.2 F 0.0
15:00 219.8 F 0.0
20:00 207.8 F 0.0
25:00 206.6 F 0.0
30:00 202.9 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

09/02/2021 03:43 PM
ACCEPTED: X
REJECTED:
Operator: U02