
Midmark M11 - v1.0.0
Total Cycles: 1555

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2022 12:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.3 F 0.0
2:00 168.0 F 3.9
4:00 184.5 F 8.4
6:00 210.2 F 10.8
8:00 225.9 F 8.1
10:00 239.2 F 13.5
12:00 238.6 F 11.0
14:00 237.5 F 9.2
16:00 251.2 F 16.1
18:00 263.1 F 23.0
20:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.2
2:00 271.5 F 28.5
3:00 272.6 F 29.3
4:00 272.0 F 28.8

Min 271.4 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 231.3 F 0.0
10:00 226.8 F 0.0
15:00 229.2 F 0.0
20:00 229.6 F 0.0
25:00 231.4 F 0.0
30:00 231.6 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:49
CYCLE COMPLETE

01/24/2022 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____