
Midmark M11 - v1.0.0
Total Cycles: 2868

Name: RITTER 1
Sterilizer ID: V2178198
02/07/2023 11:13 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.0 F 0.0
2:00 169.8 F 2.9
4:00 172.0 F 6.8
6:00 195.9 F 11.5
8:00 213.8 F 5.2
10:00 226.8 F 10.2
12:00 233.2 F 7.4
14:00 241.7 F 12.4
16:00 240.1 F 10.7
18:00 253.7 F 17.4
20:00 265.0 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.4 F 29.3
2:00 272.0 F 28.8
3:00 271.0 F 28.3
4:00 271.9 F 29.2

Min 270.9 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 217.1 F 0.0
10:00 225.5 F 0.0
15:00 231.7 F 0.0
20:00 223.2 F 0.0
25:00 225.3 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

02/07/2023 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____