
Midmark M11 - v1.0.5
Total Cycles: 924

Name: RITTER2
Sterilizer ID: V2313228
03/29/2022 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.3 F 0.0
2:00 165.2 F 0.3
4:00 192.1 F 3.3
6:00 219.5 F 8.6
8:00 230.5 F 10.7
10:00 236.3 F 10.5
12:00 234.0 F 8.3
14:00 240.5 F 10.5
16:00 248.6 F 14.8
18:00 261.3 F 21.9
20:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.7 F 29.2
2:00 271.2 F 28.5
3:00 272.6 F 29.4
4:00 271.8 F 28.8

Min 271.1 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 230.6 F 0.0
10:00 215.8 F 0.0
15:00 227.1 F 0.0
20:00 218.9 F 0.0
25:00 220.0 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 20:25
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

03/29/2022 04:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____