
Midmark M11 - v1.0.5
Total Cycles: 165

Name: RITTER2
Sterilizer ID: V2313228
07/06/2021 08:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.1 F 0.0
2:00 140.4 F 1.4
4:00 147.2 F 4.3
6:00 172.1 F 8.1
8:00 195.6 F 12.8
10:00 212.0 F 6.1
12:00 225.5 F 10.7
14:00 230.4 F 7.6
16:00 241.2 F 13.2
18:00 241.8 F 11.8
20:00 253.5 F 17.9
22:00 264.2 F 24.4
24:00 271.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 30.1
2:00 272.2 F 29.5
3:00 271.4 F 29.2
4:00 272.3 F 30.1

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.1 F 0.0
5:00 216.3 F 0.0
10:00 226.9 F 0.0
15:00 222.3 F 0.0
20:00 219.4 F 0.0
25:00 218.1 F 0.0
30:00 228.4 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 24:08
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:18
CYCLE COMPLETE

07/06/2021 09:19 AM
ACCEPTED: X
REJECTED:
Operator: U02