
Midmark M11 - v1.0.0
Total Cycles: 4773

Name: RITTER 1
Sterilizer ID: V2178198
07/29/2024 08:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.5 F 0.0
2:00 70.8 F 0.0
4:00 99.9 F 0.1
6:00 118.7 F 1.7
8:00 144.0 F 4.2
10:00 171.9 F 7.1
12:00 194.6 F 10.3
14:00 203.6 F 5.8
16:00 214.7 F 5.3
18:00 225.2 F 9.0
20:00 235.6 F 13.1
22:00 234.3 F 9.0
24:00 241.0 F 10.9
26:00 243.9 F 12.3
28:00 255.6 F 18.3
30:00 265.9 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.5 F 28.8
2:00 271.6 F 28.4
3:00 272.4 F 28.8
4:00 271.4 F 28.2

Min 271.4 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 200.7 F 0.0
10:00 213.7 F 0.0
15:00 219.8 F 0.0
20:00 222.0 F 0.0
25:00 226.8 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:47
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:56
CYCLE COMPLETE

07/29/2024 09:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____