
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006518
10:25 23/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:22
PRESSURIZING 1:24
STERILIZING 2:19
136.4 C 220kPa 2:19
136.4 C 224kPa 2:49
136.3 C 223kPa 3:19
136.2 C 221kPa 3:49
136.4 C 223kPa 4:19
135.9 C 220kPa 4:49
136.3 C 224kPa 5:19
135.9 C 220kPa 5:49
135.9 C 220kPa 5:49
VENTING 5:49
AIR DRYING 6:40
CYCLE COMPLETE 7:06

Digital Signature #
EA353E46D571E770
