

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3002

Name: RITTER 1  
Sterilizer ID: V2178198  
03/14/2023 10:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 206.0 F 0.0  
2:00 195.1 F 3.5  
4:00 209.1 F 8.5  
6:00 230.4 F 8.6  
8:00 234.8 F 10.4  
10:00 235.5 F 7.5  
12:00 241.3 F 10.9  
14:00 246.8 F 13.7  
16:00 259.9 F 20.8  
18:00 271.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.6 F 29.2  
2:00 271.8 F 28.5  
3:00 271.3 F 28.5  
4:00 272.6 F 29.2  
  
Min 271.1 F 28.1  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 217.3 F 0.0  
10:00 232.5 F 0.0  
15:00 230.4 F 0.0  
20:00 226.1 F 0.0  
25:00 218.6 F 0.0  
30:00 214.4 F 0.0  
Drying Complete

FILLING: 1:57  
HEATING: 18:40  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:50  
CYCLE COMPLETE

03/14/2023 11:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_