
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000815
11:04 25/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:56
PRESSURIZING 7:00
STERILIZING 8:31
136.0 C 220kPa 8:31
136.4 C 217kPa 9:01
136.2 C 221kPa 9:31
136.2 C 221kPa 10:01
136.4 C 223kPa 10:31
135.9 C 221kPa 11:01
136.3 C 222kPa 11:31
136.2 C 220kPa 12:01
136.2 C 220kPa 12:01
VENTING 12:01
AIR DRYING 13:04
CYCLE COMPLETE 43:04

Digital Signature #
C9F50248C6C5E1D7
