

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001473  
13:06 10/04/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:37  
PRESSURIZING 6:25  
STERILIZING 7:57  
136.0 C 220kPa 7:57  
136.4 C 220kPa 8:27  
136.5 C 222kPa 8:57  
136.2 C 221kPa 9:27  
136.4 C 222kPa 9:57  
136.4 C 222kPa 10:27  
136.2 C 221kPa 10:57  
136.3 C 222kPa 11:27  
VENTING 11:27  
AIR DRYING 12:30  
CYCLE COMPLETE 27:21  
-----

Digital Signature #  
927401D04F230581  
-----