

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4364

Name: RITTER 1  
Sterilizer ID: V2178198  
03/07/2024 11:53 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 136.3 F 0.0  
2:00 138.2 F 1.5  
4:00 145.0 F 4.4  
6:00 180.9 F 7.3  
8:00 192.2 F 11.5  
10:00 210.4 F 3.6  
12:00 218.2 F 7.5  
14:00 229.7 F 11.7  
16:00 231.2 F 7.4  
18:00 242.0 F 12.5  
20:00 239.9 F 10.4  
22:00 251.6 F 16.1  
24:00 261.9 F 22.0  
26:00 271.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 271.6 F 28.7  
2:00 272.4 F 29.0  
3:00 271.6 F 28.5  
4:00 271.2 F 28.5

Min 271.1 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 213.6 F 0.0  
10:00 221.7 F 0.0  
15:00 231.5 F 0.0  
20:00 228.4 F 0.0  
25:00 228.4 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 26:33  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:05:22  
CYCLE COMPLETE

03/07/2024 12:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_