
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003985
10:25 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:32
PRESSURIZING 1:36
STERILIZING 2:36
136.0 C 220kPa 2:36
136.4 C 224kPa 3:06
136.4 C 224kPa 3:36
136.1 C 222kPa 4:06
136.4 C 224kPa 4:36
136.4 C 224kPa 5:06
136.4 C 224kPa 5:36
136.4 C 224kPa 6:06
136.4 C 224kPa 6:06
VENTING 6:06
AIR DRYING 6:52
CYCLE COMPLETE 7:01

Digital Signature #
7F209008D45521C2
