

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000349  
11:58 19/08/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:45  
STERILIZING 3:49  
136.0 C 221kPa 3:49  
136.4 C 224kPa 4:19  
136.4 C 223kPa 4:49  
136.4 C 224kPa 5:19  
136.2 C 223kPa 5:49  
136.2 C 224kPa 6:19  
136.1 C 219kPa 6:49  
136.4 C 224kPa 7:19  
136.4 C 224kPa 7:19  
VENTING 7:19  
AIR DRYING 8:09  
CYCLE COMPLETE 8:36

-----  
Digital Signature #  
6DF38D28EBC1F81A  
-----