

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4437

Name: RITTER2  
Sterilizer ID: V2313228  
03/20/2025 08:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.1 F 0.0  
2:00 82.6 F 0.1  
4:00 128.9 F 0.4  
6:00 134.7 F 2.5  
8:00 150.2 F 5.6  
10:00 166.1 F 9.2  
12:00 186.6 F 13.1  
14:00 209.5 F 4.1  
16:00 221.9 F 7.5  
18:00 230.0 F 11.4  
20:00 229.0 F 7.3  
22:00 241.6 F 12.7  
24:00 240.2 F 11.0  
26:00 252.3 F 17.1  
28:00 262.8 F 23.4  
30:00 272.7 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.4 F 29.5  
2:00 272.5 F 29.9  
3:00 271.9 F 29.0  
4:00 272.7 F 29.7

Min 271.7 F 29.0  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.2 F 0.0  
5:00 176.7 F 0.0  
10:00 195.6 F 0.0  
15:00 197.9 F 0.0  
20:00 198.9 F 0.0  
25:00 202.6 F 0.0  
30:00 202.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 30:17  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 01:08:39  
CYCLE COMPLETE

03/20/2025 09:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_