
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000956
17:13 25/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:10
PRESSURIZING 2:23
STERILIZING 3:25
136.4 C 220kPa 3:25
136.3 C 231kPa 3:55
136.2 C 230kPa 4:25
136.2 C 230kPa 4:55
136.4 C 232kPa 5:25
136.4 C 232kPa 5:55
136.4 C 232kPa 6:25
136.3 C 231kPa 6:55
136.3 C 231kPa 6:55
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 10:40

Digital Signature #
3E9FFEA8A3926E95
