
Midmark M11 - v1.0.0
Total Cycles: 4462

Name: RITTER 1
Sterilizer ID: V2178198
04/16/2024 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.5 F 0.0
2:00 167.2 F 1.8
4:00 181.0 F 5.9
6:00 198.3 F 10.5
8:00 212.4 F 4.4
10:00 225.6 F 8.5
12:00 237.1 F 13.3
14:00 237.3 F 10.0
16:00 236.7 F 8.6
18:00 248.3 F 14.3
20:00 259.8 F 20.6
22:00 270.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.8 F 28.8
2:00 271.5 F 28.1
3:00 272.8 F 29.0
4:00 272.0 F 28.4

Min 271.4 F 28.1
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 226.1 F 0.0
10:00 234.0 F 0.0
15:00 228.0 F 0.0
20:00 223.1 F 0.0
25:00 224.7 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:52
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:01:37
CYCLE COMPLETE

04/16/2024 10:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____