
Midmark M11 - v1.0.0
Total Cycles: 3599

Name: RITTER 1
Sterilizer ID: V2178198
08/01/2023 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 212.6 F 0.0
2:00 207.7 F 3.7
4:00 212.4 F 9.1
6:00 223.0 F 6.6
8:00 234.4 F 10.6
10:00 235.0 F 7.9
12:00 241.5 F 10.7
14:00 247.0 F 13.6
16:00 260.0 F 20.7
18:00 271.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.9
2:00 271.5 F 28.3
3:00 272.5 F 29.0
4:00 271.7 F 28.2

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 223.8 F 0.0
10:00 233.4 F 0.0
15:00 227.4 F 0.0
20:00 227.4 F 0.0
25:00 230.9 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 18:39
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

08/01/2023 11:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____