

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.1uS / 1.3ppm  
CYCLE NUMBER 003690  
9:59 26/05/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:09  
PRESSURIZING 2:13  
STERILIZING 3:13  
136.0 C 220kPa 3:13  
136.1 C 222kPa 3:43  
136.3 C 223kPa 4:13  
136.4 C 223kPa 4:43  
136.4 C 223kPa 5:13  
136.4 C 224kPa 5:43  
136.4 C 223kPa 6:13  
136.1 C 221kPa 6:43  
136.1 C 221kPa 6:43  
VENTING 6:43  
AIR DRYING 8:07  
CYCLE COMPLETE 8:22

-----  
Digital Signature #  
4E4012128ABC5C96  
-----