
Midmark M11 - v1.0.0
Total Cycles: 4020

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2023 09:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 73.8 F 0.1
4:00 84.2 F 1.0
6:00 113.7 F 3.0
8:00 141.4 F 5.5
10:00 164.4 F 8.4
12:00 182.8 F 11.7
14:00 202.8 F 2.1
16:00 209.3 F 5.3
18:00 220.0 F 8.6
20:00 230.0 F 12.2
22:00 229.9 F 7.6
24:00 241.0 F 12.5
26:00 239.0 F 10.1
28:00 250.7 F 15.8
30:00 261.0 F 21.5
32:00 270.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.1
2:00 271.8 F 28.5
3:00 271.3 F 28.5
4:00 272.6 F 29.1

Min 271.0 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 222.5 F 0.0
10:00 219.2 F 0.0
15:00 227.5 F 0.0
20:00 232.2 F 0.0
25:00 230.4 F 0.0
30:00 228.1 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:53
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:11:19
CYCLE COMPLETE

11/15/2023 10:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____