

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000513  
7:04 28/09/2020

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:26  
PRESSURIZING 3:27  
STERILIZING 4:36  
136.4 C 223kPa 4:36  
136.3 C 220kPa 5:06  
136.1 C 221kPa 5:36  
136.4 C 223kPa 6:06  
136.3 C 223kPa 6:36  
136.4 C 223kPa 7:06  
136.3 C 223kPa 7:36  
136.3 C 222kPa 8:06  
136.3 C 222kPa 8:06  
VENTING 8:07  
AIR DRYING 8:59  
CYCLE COMPLETE 9:18

-----  
Digital Signature #  
D1AEF859FA8E760A  
-----