

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2327

Name: RITTER2  
Sterilizer ID: V2313228  
05/02/2023 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.8 F 0.0  
2:00 187.8 F 3.6  
4:00 195.3 F 8.5  
6:00 215.5 F 10.1  
8:00 228.3 F 9.7  
10:00 233.7 F 9.1  
12:00 242.7 F 13.2  
14:00 242.7 F 12.3  
16:00 255.5 F 19.0  
18:00 266.8 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.5 F 29.9  
2:00 271.5 F 29.2  
3:00 272.6 F 30.0  
4:00 272.0 F 29.5  
  
Min 271.3 F 29.1  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.4 F 0.0  
5:00 213.5 F 0.0  
10:00 231.4 F 0.0  
15:00 234.1 F 0.0  
20:00 221.0 F 0.0  
25:00 220.8 F 0.0  
30:00 216.8 F 0.0  
Drying Complete

FILLING: 2:02  
HEATING: 19:31  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

05/02/2023 03:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_