
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001783
12:18 22/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:21
STERILIZING 3:25
136.0 C 220kPa 3:25
136.4 C 225kPa 3:55
136.4 C 224kPa 4:25
136.2 C 222kPa 4:55
136.2 C 223kPa 5:25
136.4 C 224kPa 5:55
136.4 C 225kPa 6:25
136.4 C 222kPa 6:55
136.4 C 222kPa 6:55
VENTING 6:55
AIR DRYING 7:47
CYCLE COMPLETE 17:08

Digital Signature #
423640D4E1295F5F
