

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4484

Name: RITTER2  
Sterilizer ID: V2313228  
04/03/2025 01:26 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.7 F 0.0  
2:00 141.6 F 2.0  
4:00 149.6 F 5.5  
6:00 185.3 F 9.8  
8:00 197.2 F 7.2  
10:00 214.0 F 7.6  
12:00 227.6 F 12.0  
14:00 233.0 F 8.7  
16:00 239.9 F 11.4  
18:00 243.8 F 12.9  
20:00 255.4 F 19.0  
22:00 265.5 F 25.4  
24:00 271.3 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 30.1  
1:00 272.5 F 30.6  
2:00 271.5 F 30.2  
3:00 272.1 F 30.6  
4:00 271.3 F 30.1

Min 271.3 F 30.0  
Max 272.5 F 30.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.5 F 0.0  
5:00 194.3 F 0.0  
10:00 183.4 F 0.0  
15:00 187.9 F 0.0  
20:00 188.4 F 0.0  
25:00 189.0 F 0.0  
30:00 189.3 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 24:40  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:03:36  
CYCLE COMPLETE

04/03/2025 02:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_