
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002287
15:41 18/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:33
STERILIZING 3:37
136.5 C 224kPa 3:37
136.2 C 224kPa 4:07
136.4 C 226kPa 4:37
136.3 C 225kPa 5:07
136.4 C 226kPa 5:37
136.3 C 225kPa 6:07
136.2 C 223kPa 6:37
136.3 C 225kPa 7:07
136.3 C 225kPa 7:07
VENTING 7:07
AIR DRYING 8:01
CYCLE COMPLETE 9:11

Digital Signature #
6D9181B2FDFAADA1
