
Midmark M11 - v1.0.5
Total Cycles: 4153

Name: RITTER2
Sterilizer ID: V2313228
11/22/2024 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.6 F 0.0
2:00 141.6 F 0.0
4:00 143.0 F 0.3
6:00 178.3 F 3.1
8:00 197.4 F 6.6
10:00 214.2 F 10.3
12:00 221.1 F 8.1
14:00 226.8 F 9.2
16:00 235.3 F 11.7
18:00 239.7 F 11.5
20:00 237.8 F 10.1
22:00 250.4 F 16.2
24:00 261.2 F 22.3
26:00 271.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 30.0
2:00 271.9 F 29.3
3:00 271.4 F 29.4
4:00 272.3 F 30.0

Min 271.2 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 209.6 F 0.0
10:00 221.8 F 0.0
15:00 215.3 F 0.0
20:00 208.2 F 0.0
25:00 212.3 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 26:46
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:25
CYCLE COMPLETE

11/22/2024 12:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____