
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007564
10:00 18/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:41
STERILIZING 4:38
136.3 C 220kPa 4:38
135.9 C 221kPa 5:08
136.3 C 223kPa 5:38
136.4 C 224kPa 6:08
136.1 C 221kPa 6:38
136.4 C 224kPa 7:08
136.2 C 217kPa 7:38
136.3 C 223kPa 8:08
136.3 C 223kPa 8:08
VENTING 8:08
AIR DRYING 9:00
CYCLE COMPLETE 21:23

Digital Signature #
55F872C0EFB1FC03
