
Midmark M11 - v1.0.0
Total Cycles: 3145

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2023 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.6 F 0.0
2:00 195.3 F 0.0
4:00 209.0 F 1.4
6:00 222.9 F 5.9
8:00 234.1 F 10.4
10:00 237.0 F 8.2
12:00 244.4 F 13.1
14:00 243.4 F 12.0
16:00 242.5 F 11.3
18:00 255.7 F 18.0
20:00 267.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 272.5 F 28.7
2:00 272.0 F 28.1
3:00 272.0 F 28.4
4:00 272.6 F 28.5

Min 271.4 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 210.0 F 0.0
10:00 221.1 F 0.0
15:00 223.4 F 0.0
20:00 222.2 F 0.0
25:00 220.2 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:21
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

04/21/2023 01:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____