
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001348
13:36 15/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:08
PRESSURIZING 4:09
STERILIZING 5:16
136.0 C 220kPa 5:16
136.4 C 225kPa 5:46
135.9 C 222kPa 6:16
136.4 C 225kPa 6:46
136.4 C 225kPa 7:16
136.4 C 223kPa 7:46
136.1 C 222kPa 8:16
136.4 C 225kPa 8:46
136.4 C 225kPa 8:46
VENTING 8:46
AIR DRYING 9:40
CYCLE COMPLETE 24:40

Digital Signature #
E2E0EAE6B66B96ED
