

-----  
Midmark M11 - v1.0.5  
Total Cycles: 303

Name: RITTER2  
Sterilizer ID: V2313228  
08/23/2021 01:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.2 F 0.0  
2:00 161.1 F 3.2  
4:00 184.0 F 7.6  
6:00 209.7 F 13.2  
8:00 222.3 F 8.5  
10:00 236.2 F 11.1  
12:00 240.6 F 11.8  
14:00 239.4 F 10.7  
16:00 253.1 F 17.5  
18:00 264.9 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.0  
1:00 272.2 F 29.3  
2:00 271.7 F 29.7  
3:00 271.9 F 29.1  
4:00 272.5 F 30.2  
  
Min 271.0 F 28.8  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.8 F 0.0  
5:00 204.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:39  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 5:03  
TOTAL CYCLE: 00:33:33

08/23/2021 02:10 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02