

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000441  
12:25 28/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:43  
STERILIZING 3:40  
136.4 C 220kPa 3:40  
136.4 C 233kPa 4:10  
136.1 C 231kPa 4:40  
136.4 C 233kPa 5:10  
136.3 C 232kPa 5:40  
136.3 C 233kPa 6:10  
136.1 C 231kPa 6:40  
136.4 C 231kPa 7:10  
136.4 C 231kPa 7:10  
VENTING 7:10  
AIR DRYING 8:03  
CYCLE COMPLETE 8:10  
-----

Digital Signature #  
09DCD2A34410A7C6  
-----