

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3987

Name: RITTER 1  
Sterilizer ID: V2178198  
11/07/2023 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 85.0 F 0.2  
4:00 129.4 F 1.2  
6:00 125.7 F 3.4  
8:00 145.8 F 6.5  
10:00 171.3 F 10.2  
12:00 197.4 F 6.6  
14:00 213.9 F 5.0  
16:00 224.9 F 8.4  
18:00 233.4 F 12.3  
20:00 231.2 F 7.6  
22:00 242.7 F 12.6  
24:00 239.3 F 10.2  
26:00 251.7 F 16.4  
28:00 262.9 F 22.8  
30:00 272.3 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.4  
1:00 272.5 F 29.2  
2:00 271.8 F 28.4  
3:00 272.6 F 29.2  
4:00 272.2 F 28.6

Min 271.3 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.1 F 0.0  
5:00 238.8 F 0.0  
10:00 224.4 F 0.0  
15:00 231.1 F 0.0  
20:00 220.9 F 0.0  
25:00 213.6 F 0.0  
30:00 207.4 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 30:06  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:08:17  
CYCLE COMPLETE

11/07/2023 10:12 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_