

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 000766  
8:33 09/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:56  
PRESSURIZING 3:57  
STERILIZING 5:07  
136.0 C 220kPa 5:07  
136.4 C 224kPa 5:37  
136.2 C 224kPa 6:07  
136.4 C 224kPa 6:37  
136.3 C 223kPa 7:07  
136.3 C 224kPa 7:37  
136.3 C 222kPa 8:07  
136.4 C 224kPa 8:37  
136.4 C 224kPa 8:37  
VENTING 8:37  
AIR DRYING 9:32  
CYCLE COMPLETE 9:37

-----  
Digital Signature #  
20A6C4FC64CA770D  
-----