
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000529
9:51 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:15
PRESSURIZING 3:16
STERILIZING 4:22
136.0 C 220kPa 4:22
136.4 C 225kPa 4:52
136.3 C 223kPa 5:22
136.4 C 225kPa 5:52
136.3 C 224kPa 6:22
136.3 C 224kPa 6:52
136.4 C 223kPa 7:22
136.4 C 223kPa 7:52
136.4 C 223kPa 7:52
VENTING 7:52
AIR DRYING 8:45
CYCLE COMPLETE 8:50

Digital Signature #
C222E3B13A8BD06D
