

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001477  
13:58 25/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:06  
STERILIZING 4:06  
136.0 C 220kPa 4:06  
136.4 C 224kPa 4:36  
136.4 C 225kPa 5:06  
136.3 C 225kPa 5:36  
136.4 C 226kPa 6:06  
136.4 C 225kPa 6:36  
136.4 C 225kPa 7:06  
136.4 C 225kPa 7:36  
136.4 C 225kPa 7:36  
VENTING 7:36  
AIR DRYING 8:29  
CYCLE COMPLETE 9:05  
-----

Digital Signature #  
ABA15CB82A863398  
-----